



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

FASI, Inc.

22 Pine Street, Suite #102, Bristol, CT 06010

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Dimensional Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

February 26, 2010

Issue Date:

December 27, 2023

Expiration Date:

March 31, 2026

Accreditation No.:

67316

Certificate No.:

L23-937

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjllabs.com



Certificate of Accreditation: Supplement

FASI, Inc.

22 Pine Street, Suite #102, Bristol, CT 06010
Contact Name: Mr. John Youngs Phone: 937-898-4300

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED |
|--|---|--|--|---|
| Laser Micrometer ^{FO} | 0.004 in to 1 in | 18 μ in | Class XXX Gauge Pins | CP-9 |
| | 1 in to 2 in | 34 μ in | | |
| | 2 in to 4 in | 82 μ in | | |
| | 4 in to 5.5 in | 310 μ in | | |
| | 5.5 in to 7.5 in | 520 μ in | | |
| | 7.5 in to 12 in | 1 900 μ in | | |
| Proton Products Speed & Length Gauge ^{FO} | 10 m to 10 000 m | (0.000 5 + 0.000 6L) m | Proton Encoder Wheel 5 in | CP-20 |

- The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
- The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
- The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations.
- The term L represents length in inches or millimeters as appropriate to the uncertainty statement
- Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
- This organization maintains satellite organization(s) where no key activities are performed. The accredited corporate site with the above address is also accredited for satellite site(s). Only one certificate and scope of accreditation is issued with the corporate organization's address. Reports are issued from the corporate address only.
- This organization maintains the following satellite sites: 521 Byers Road, Suite #107, Miamisburg, OH 45342; 4008 Medlin Road, Monroe, NC 28112; 107 N. US Highway 169, Smithville, MO 64089